

**CHIEF DISTRICT MEDICAL & PUBLIC HEALTH OFFICER, MALKANGIRI**

No. 10427/ XIV-P(D&C)-16/20-, dated Malkangiri the 29.10.2020

**CORRIGENDUM**

After recommendation of the tender committee after pre-bid conference for procurement of Medical Equipments under SCA schemes vides Bid Reference No.9985 Dt.15.10.2020 some alternation of the specifications of some items has been made. The bidders can be downloaded the modified specifications directly from the website available at [www.malkangiri.nic.in](http://www.malkangiri.nic.in) and submit their tender accordingly. There is no change of date and time schedule for receipt of the tender document and opening technical Bid.

**Sd/- P.K.Nanda**

**Chief District Medical & Public Health Officer, Malkangiri**

**CHIEF DISTRICT MEDICAL & PUBLIC HEALTH OFFICER, MALKANGIRI**

No.10427/XIV-P(D&C)-16/2020, dated, Malkangiri the 27.10.2020

**CORRIGENDUM**

After pre-bid conference for procurement of Medical Equipments under SCA schemes held on 27.10.2020 at 05.30 P.M. in response to Bid Reference No.9985 Dt.15.10.2020 of CDM&PHO, Malkangiri the following modification of the Specification of some items of the Original Tender has been made. Hence the bidders have to download the modified Specifications directly from the website available at [www.malkangiri.nic.in](http://www.malkangiri.nic.in) and submit their tender accordingly. There is no change of time schedule for receipt of the tender and opening of the technical Bid.

The details changes/modifications are given below.

Sl. No	Tender Specification	Quarry raised by bidders	Remarks
<b>2.</b>	<b>300 mA HF X-ray machine</b>		
1	6. ENVIRONMENTAL AND DEPARTMENTAL CONSIDERATIONS 6.1 Atmosphere/Ambiance (air conditioning, humidity, dust ...) 1) Operating condition: Capable of operating continuously in ambient temperature of 5 to 50 deg C and relative humidity of 15 to 80% in ideal circumstances. 2) Storage condition: Capable of being stored continuously in ambient temperature of 0 to 50 deg C and relative humidity of 15 to 90%.	Environment and Departmental Consideration. 6.1. Atmosphere / Ambiance (air conditioning humidity, dust...) 1). Operating condition: Capable of operating continuously in ambient temperature of 5 to 50 deg C and relative humidity of 15 to 80 % in deal circumstances. 2) STORAGE CONDITION: Capable of being stored continuously in ambient temperature of 0 to 50 deg C and relative humidity of 15 to 90%.	Amended: Operating Condition: Temperature 10 to 40 degree Centigrade  Humidity: 30 to 80% Storage Conditions: Temperature: 0-40 degree Centigrade  Humidity: maximum 90% at 30 degree centigrade
2	7. STANDARDS AND SAFETY 7.1 Certificates (pre-market, sanitary); Performance and safety standards (specific to the device type); Local and/or international 1. Should be FDA/European CE/BIS approved product. 2. Manufacturer and Supplier should have ISO 13485 certification for quality standards.	7. Standards and Safety 1. Should be FDA/ European CE/BIS approved product.	2.Standards and Safety: Should be USFDA/BIS / European CE (Notified Body)

Sl. No	Tender Specification	Quarry raised by bidders	Remarks
2.	<u>300 mA HF X-ray machine</u>		
3	Delivery: Within 60 days from issue of the purchase order	60 days from the date of confirmation of e-LORA from the hospital authorities.	No Change
4	<u>Comprehensive warranty</u> including all spares, maintenance etc. for a period <i>3(three) years</i> from the date of installation & commissioning and <i>3(three) years</i> CMC after warranty period.	Kindly clarify how many years of CMC we will follow.	Only 3 years Warranty, NO CMC (CMC for 5years/3years may be ignored) as the state has implemented Centralize CMC programme.

Sd/- P.K.Nanda  
Chief District Medical & Public Health Officer, Malkangiri