

373

By e-Mail- orimal@nic.in

CHIEF DISTRICT MEDICAL & PUBLIC HEALTH OFFICER, MALKANGIRI

No. 12012 //XIV-P (D&C)-16/2021, dated, Malkangiri the 1st Nov 2021

To

The District Informatics Officer,
NIC, Malkangiri.

Sub: - Publication of advertisement of Corrigendum Notice for **SUPPLY & INSTALLATION MEDICAL EQUIPMENT FOR MALKANGIRI DISTRICT (UNDER SCA SCHEMES)** in response to this Office Bid Reference No.11703 Dt.22.10.2021.

Ref:- This Office Letter No.11741 Dt.22.10.2021.

Sir,

In inviting a reference to this office Letter No.11741 Dt.22.10.2021, enclosed please find herewith the advertisement of Corrigendum Notice for **SUPPLY & INSTALLATION OF MEDICAL EQUIPMENT FOR MALKANGIRI DISTRICT (UNDER SCA SCHEMES)** both hard copy and soft copy for publication in Malkangiri website www.malkangiri.nic.in. The last of receipt of Tender Documents is 06.11.2021 up to 05.00 PM and opening of Technical Bid is 06.11.2021, 05.30 PM.

Yours Faithfully,


30/10/21

Chief District Medical & Public Health Officer, Malkangiri


30/10/21

CHIEF DISTRICT MEDICAL & PUBLIC HEALTH OFFICER, MALKANGIRI

No. 12011/XIV-P (D&C)-16/2021, dated, Malkangiri the 1st Nov'2021

CORRIGENDUM

After pre-bid conference for procurement of Medical Equipments under SCA schemes held on 29.10.2021 at 11.00 AM in response to Bid Reference No.11703 Dt.22.10.2021 of CDM&PHO, Malkangiri the following modification of the Specification of some items of the Original Tender has been made.

The details changes/modifications are given below.

Sl. No	Tender Specification	Quarry raised by bidders	Remarks
Name of the Item: Remote & Table mounted General Operating Table			
1	Table Features :- Type of OT table Actuator Electro-hydraulic	Electro-Manual	Amendment : "Electro Hydraulic" amended as "Electro Manual"
2	Table Features:- Table top shall be X-ray compactable and stain Resistant-Yes		Amendment: "Deleted" from Original Specification

Sd/-P.K.Nanda

Chief District Medical & Public Health Officer, Malkangiri